

HUMAN SERUM QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

Date of Manufacture:	2020-12-29 CALIBRATION SERUM LEVEL 2 (CAL 2) CAL2350		
Product Description:			
Catalogue Number:			
Batch Number:	1528UN		
Expiry Date:	2023-06-28		

Analyte	Lot No.	Target	arget Range		Result	Units
Albumin	1528UN	41.0	N/A	N/A	40.3	g/l
Total Protein	1528UN	57.8	N/A	N/A	57.7	g/l

Human source material from which this product was derived has been tested at donor level for HIV1 & HIV 2, Syphilis, HbsAG, and Antibody to HCV and found to be negative. However, since no method can offer complete assurance as to the absence of infectious agents this material should be handled with the same precautions recommended for a human serum or blood specimen.

Approved By: J. williamson Date: 22-0ct-21

Quality Control Representative