

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2020-10-29
 KIT DESCRIPTION: IMMUNOASSAY SPECIALITY II - LEVEL 2
 CATALOGUE NO: IAS3118
 LOT NUMBERS: 2018EC
 EXPIRY DATE: 2022-10-28
 SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Gastrin (pmol/L)	2018EC	37.8 - 63.0	50.4	51.5	N/A
				49.1	
Calcitonin (pg/ml)	2018EC	85.6 - 142	114	121	N/A
				120	

QC TESTED BY: Charlene Junkin
 DATE: 26-Feb-21
 RESULT (PASS/FAIL): PASS
 QC APPROVED BY: *C. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
 T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAC Man R J Fitzgerald, W Fitzgerald OBE P P Kelly BSc FCA
 Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738 VAT Number GB 353 0304 00.