



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

| | |
|-----------------------------|------------------------------|
| DATE OF MANUFACTURE: | 2019-06-29 |
| KIT DESCRIPTION: | LIPID CONTROL LEVEL 3 |
| CATALOGUE NO: | LE2670 |
| BATCH NO: | 2727CH |
| EXPIRY DATE: | 2022-06-28 |

| SAMPLE | ANALYTE | TARGET | RANGE | RESULT | UNITS |
|---------------|----------------|---------------|------------------|---------------|---------------|
| 2727CH | DHDL | 1.75 | 1.49-2.01 | 1.77 | mmol/L |
| 2727CH | DLDL | 5.14 | 4.37-5.91 | 5.31 | mmol/L |

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRAE MInst R J Fitzgerald W Fitzgerald OBE R F Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.