

RANDOX


CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2020-01-29
KIT DESCRIPTION: Maternal Screening Control 3
CATALOGUE NO: MSS5026
LOT NUMBERS: 7703MS
EXPIRY DATE: 2023-01-28
SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

TEST CONTROLS OR VIALS

RESULTS FOR TEST CONTROLS OR VIALS	LOT NO	RANGE	TARGET	RESULT	PRECISION (if applicable)
ANALYTE					
Free Beta hCG (mU/mL)	7703MS	176-294	235	269.0	
				257.0	
PAPP-A (U/L)	7703MS	10.4-17.3	13.80	15.5	
				15.6	
PASS or FAIL	PASS				

QC TESTED BY: Emer Ni Bhraonain
DATE: 7-May 20
RESULT (PASS/FAIL): PASS
QC APPROVED BY: 

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016