

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2021-07-29
KIT DESCRIPTION: TUMOUR MARKER CONTROL - LEVEL 3
CATALOGUE NO: TU5003
LOT NUMBERS: 333TU
EXPIRY DATE: 2023-07-28
SYSTEM OF ANALYSIS: ROCHE COBAS e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
CEA (µg/l)	333TU	32.8-49.2	41.0	42.0	N/A
				41.4	
PSA Free (µg/l)	333TU	19.7-32.9	26.3	27.0	N/A
				27.0	

QC TESTED BY: Davina Graham
DATE: 12-Aug-21
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *A. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number: NI15738, VAT Number: GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016